

Subject: FW: CS-16055 Product malfunction complaint; CTC Analytics GC PAL

From: Toni EBERWEIN <support@ctc.ch>

Sent: 05 December 2024 17:05

To: Oasis Lab <oasislab@iisc.ac.in>

Subject: CS-16055 Product malfunction complaint; CTC Analytics GC PAL

External Email

Reply above this line.

Dear Ritvik

Sure, see below the information. From the pictures you sent, I believe it is rather the display itself. Please note that also the complete Terminal is available.

Best regards
Toni

PN: Cbl SS8J-700

Spiral Cable 8-pole Shielded, RJ45 Connectors, 700 mm long


(Connection between Terminal and PAL)

(TMO PN not available)



TMO PN: **00950-01-00396**
LCD, RPL, CONTROLLER, OPEN A/S, RoHS

CTC PN: **Kit LCD Repl**
Replacement LCD for PAL Control Terminal, consisting of:
LCD-Module 128x128 mm
Flat Cable
Screws, Spacers, Star Washers



PN: PAL TermCTC-xt
Control Terminal for PAL-xt System

TMO PN: BRE0015298



[View request](#) · [Turn off this request's notifications](#)

This is shared with ritvik@iisc.ac.in, pghosh@iisc.ac.in, rachitaghosh@iisc.ac.in, and 6 other people.

Customer Support

CTC Analytics AG
Industriestrasse 20, 4222 Zwingen
Switzerland

www.ctc.ch · www.palsystem.com
